

ABSTRACT OF THE DISCLOSURE

The properties of features formed in a substrate are measured. Interferometric illumination is used to illuminate regions of a substrate so that the features of interest occupy a greater proportion of the illuminated area. The signal-to-noise ratio of the measurement signal is therefore increased, and the sensitivity of the measurement is thus improved.

449142_1.DOC